


DB 100



Technical features

-: prepolarised electret condenser.
.....: 20 mV/Pa .
-: IEC 61672-1 Class 2 /
IEC 60651 Class 2 / IEC 60804 Class 2
.....: L_A and L_{Aeq}
.....: L_{AFmax} , L_{AFmin} , L_{ASmax} , L_{ASmin}
.....: A
.....: 30-130 dB
.....: slow, fast
Data integration time for L_{Aeq}: 1 ~ 15
.....
.....: 128 x 64
.....: 가
.....: 0,1 dB
.....
.....: 30 - 130 dB
.....: 94 dB
.....: 1000 Hz
-: 95 % RH max.
.....: 0 ~ +50
.....: - 10 ~ +50
.....: 30 ~ 90% RH
.....: 65%RH, 40
.....: Class 2
Standards.....: IEC 61672-1 / IEC 61651 / IEC 60804
.....: 89/336/CEE
-: 3AAA
(Rq: rechargeable batteries must not be recharged inside
the instrument)
(20): 30 ()

Output

 **DO NOT PLUG USB cable.** The output is not USB compatible, the plug is maintenance- and optional accessory-specific.

Description

- DB 100
-
- 가
- L_A
FAST / SLOW 가
- (L_A) dBA, L_{Amax} , L_{Amin}
- L_{Aeq}
가
() dBA
(min), (second)

CTL 100



Automatic check of sound level meter

Principle of automatic check

Initial check

To be carried out at the delivery, when instrument is new and calibrated (laboratory or manufacturer) or after periodic calibration procedure, or after repair.

Frequent check

To be carried out **BEFORE** :

- each measurement dataset

To be carried out **AFTER** :

- an impact applied on the instrument,
- storage in extreme environment (high temperature, wet environment etc...)
- a long period of storage

Working principle

CTL100 gives a stable acoustic signal **90 dB at 1000Hz**, automatically delivered once plugged to the sonometer. The user shall write down the **LA value**, fast (F) or slow (S) displayed on the sound level meter.

The sound level meter value and the CTL100 reference value must not exceed **90 dB ± 2dB** difference.

In case of greater difference, the sound level meter shall be returned to Customer Service Department.

Note : The sound level meter can not be calibrated with the CTL100. An acoustic calibrator must be used to calibrate sound level meters or the instrument can be sent to specialized laboratories or Customer Service Department. CTL100 works only for DB100.

Operating procedure :

- 1. Connection** → Connect CTL100 to the sound level meter plug (on the left side)
- 2. Place CTL100** → Place the DB100 inside CTL100 cavity until you reach the bottom
- 3. Check** → Monitor the sound level value displayed on the sound level meter until measurement stabilization and write down the value
- 4. Compare** → Compare displayed value : **90,5 dB*** with the written value : **90,2 dB***
- 5. Decision** →
a- If the difference is **smaller** than ±2 dB :
→ **The sound level meter is ok**
b- If the difference is **greater** than ±2 dB :
→ **The sound level meter requires overhaul**

*example



*Sound level meter supplied separately

Presentation

The automatic check consists in comparing sound level meter value with level produced by CTL100. The principle allows to periodically check sound level meter performance, especially the microphone performance which is the sensing element of the instrument.

CTL100 can not replace an acoustic calibrator which must be used for sound level meter calibration.

Technical features

Emission

Frequency.....1000 Hz ± 5%
Level.....90 dB ± 1dB
Stability.....< 0.5 dB

Automatic power supply

When being connected to the sound level meter

Environment

Operating temperature.....from +5 °C to + 40 °C
Pressure.....1013 hPa ± 10%
Storage relative humidity..... 80 % RH max.
CE labelling.....As per 89/336/CEE guideline

Dimensions

Dimensions (Without cable).....140 x 28 x 25 mm
Weight.....50 g